



Molecular  
Dimensions

A CALIBRE SCIENTIFIC COMPANY

# Data Sheet


ANTcryo™  
Ideal for Spa

## Cryo-EM Grids

### Key Info

ANTcryo™ holey support film (Amorphous Nickel Titanium alloy Grid). A new and sophisticated solution for cryo-EM single-particle sample preparation. Users of the ANTcryo™ Grid have published more than 20 papers on high-index Journal in last 3 years, 6 of which were published on CNS [2-13].

ANTcryo™ has extraordinary benefits and is widely recognised by cryo-EM community.



### Advantages

#### 1 - Improving the resolution of reconstruction

ANTcryo™ Grids offer increased resolution compared to carbon grids due to their superior conductivity, resulting in reduced electron charging effects and less beam-induced motion (BIM) during cryo-EM imaging. This improvement enables the achievement of higher-resolution structures, with user experiments demonstrating an improvement of more than 0.2 Å in resolution.

#### 2 - Improving the rate of sample preparation

ANTcryo™ support film exhibits an adsorption force nearly 20 times lower than traditional carbon film, thereby enhancing the particle inlet rate and reducing particle aggregation within the pore. This improvement significantly enhances the success rate of sample preparation by improving particle dispersion.

#### 3 - Improving the efficiency of data collection

With the contribution of higher particle density, users need much less EM time for data collection for the same number of particles. Similar to carbon films, ANTcryo™ Grids are made of amorphous material that is suitable for coma alignment. Unlike gold films, ANTA film's amorphous structure allows routine TEM alignment like carbon films (defocus estimation, stigmatism adjustment, coma-free alignment). This eliminates challenges associated with gold's polycrystalline nature, simplifying cryo-EM workflows and saving valuable research time.

ANTcryo™ film




2.4 Å

Carbon film



2.6 Å

ANTcryo™ film



Carbon film






Figure 1: Cryo-EM photo of apoferritin Compared with traditional carbon membrane, ANTcryo™ support membrane has higher particle density, less overlapping of particle dispersion, and shorter data collection time; the particle boundary is clear and the resolution obtained much higher.

ANTcryo™ film



Carbon film





**Molecular  
Dimensions**

A CALIBRE SCIENTIFIC COMPANY

# Data Sheet

ANTcryo™

The Best Grid for Spa


## Cryo-EM Grids

### Cases

#### Case 1

On February 8, 2022, Science Online published a groundbreaking paper by the Xu Huaqiang/Yin Wanchao team and the Jimin Kexin/Deng YuJun's team from the Shanghai Institute of Materia Medica, Chinese Academy of Sciences. Titled "Structures of the Omicron Spike Trimer with ACE2 and an Anti-Omicron Antibody," this study showcased the remarkable capabilities of ANTCryo grids compared to UltrAuFoil. Using ANTCryo grids, the team achieved an exceptional resolution of 2.69 Å, significantly outperforming the 2.77 Å resolution obtained with UltrAuFoil. These findings underscore ANTCryo's superior performance in revealing the intricate structures of the Omicron mutant strain of the novel coronavirus spike protein when bound to the specific therapeutic antibody JMB2002 and the receptor ACE2.

#### ANTcryo™ Film



EMDB: 32684  
PDB: 7WPE 2.69 Å

#### UltrAuFoil film




EMDB: 32680  
PDB: 7WPA 2.77 Å

#### Case 2


On March 28th 2022, the team led by Brian K. Kobilka at Stanford University published "Calcineurin Facilitates Cryo-EM Structure Determination of a Family A GPCR" on BioRxiv. In this study, cryo-EM sample preparation was conducted using ANTCryo and Quantifoil grids. The team examined various states of beta2 adrenergic receptors (β2AR) and achieved high resolutions of 3.49 Å using ANTCryo compared with 3.95 Å using Quantifoil.

#### ANTcryo™ Film



3.49 Å

#### UltrAuFoil film




3.95 Å

#### Case 3


In February 2020, the paper titled "Activation and Signaling Mechanism Revealed by Cannabinoid Receptor-Gi Complex Structures" was published by Liu Zhijie's team at ShanghaiTech University in the journal "Cell". This study analyzed the three-dimensional cryo-EM structures of CB1 and CB2 bound to G protein molecules. The cryo-EM sample preparation method utilized in this research employed the ANTCryo grid (formerly known as CryoMatrix), achieving final resolutions of 3.0 Å and 2.9 Å for CB1 and CB2, respectively.

#### CB1



EMDB: 0745  
PDB: 6KPG

#### CB2



EMDB: 0744  
PDB: 6KPF





Molecular  
Dimensions

A CALIBRE SCIENTIFIC COMPANY

# Data Sheet

ANTcryo™

The Best Grid for Spa

## Cryo-EM Grids

### References

- [1]. 2020. Prog Biophys Mol Biol, Huang, X., et al., Amorphous nickel titanium alloy film: A new choice for cryo electron microscopy sample preparation
- [2]. 2021. Nat Commun, Kang, H., et al., Structural basis for recognition and regulation of arenavirus polymerase L by Z protein.
- [3]. 2022. bioRxiv, Li, Z., et al., Cryo-EM structures reveal the activation and substrate recognition mechanism of human enteropeptidase.
- [4]. 2022. Cell Discov, Wang, A., et al., Structure of infective Getah virus at 2.8 Å resolution determined by cryo-electron microscopy.
- [5]. 2022. Nat Commun, Wang, Q., et al., Structural basis of the ligand binding and signaling mechanism of melatonin receptors.
- [6]. 2022. bioRxiv, Xu, J., et al., Calcineurin-fusion facilitates Cryo-EM Structure Determination of a Family A GPCR.
- [7]. 2022. Nat Struct Mol Biol, Zhang, H., et al., Structure of human glycosylphosphatidylinositol transamidase.
- [8]. 2021. Proc Natl Acad Sci U S A, Zhou, X., et al., Structural basis for ligand binding modes of CTP synthase.
- [9]. 2020. Cell, Hua, T., et al., Activation and Signaling Mechanism Revealed by Cannabinoid Receptor-Gi Complex Structures.
- [10]. 2021. Nature, Lin, S., et al., Structures of Gi-bound metabotropic glutamate receptors mGlu2 and mGlu4.
- [11]. 2020. Science, Qiao, A., et al., Structural basis of Gs and Gi recognition by the human glucagon receptor.
- [12]. 2020. Nature, Ruan, Z., et al., Structures of human pannexin 1 reveal ion pathways and mechanism of gating.
- [13]. 2022. Science, Yin, W., et al., Structures of the Omicron spike trimer with ACE2 and an anti-Omicron antibody.

### Ordering Information

Catalogue Number	Grid Type	Meshes	Hole Size	Spacing
MDL-M01	Au	300	1.2 µm	1.3 µm
MDL-M02	Cu	300	1.2 µm	1.3 µm
MDL-M03	Au	400	1.2 µm	1.3 µm
MDL-M04	Au	300	2.0 µm	1.0 µm

1. Grid customisation with further variations of formats are available.
2. Film Thickness  $22 \pm 3$  nm.
3. Quantity: 50 ea/pack.

