

Agar Scientific Ltd Unit 7, M11 Business Link Parsonage Lane, Stansted Essex, UK CM24 8GF t: +44 (0)1279 215 506 f: +44 (0)1279 813 105 e: sales@agarscientific.com w: www.agarscientific.com

RoHS

COMPLIANT

## **Strain Gage Adhesive**



# OTHER ACCESSORIES USED IN AN M-BOND 610 INSTALLATION:

of gluelines <0.0002 in [0.005mm]. Extremely thin, hard, void-

free gluelines minimize creep, hysteresis, and linearity

problems. Cure must begin within four hours of application.

- □ CSM Degreaser or GC-6 Isopropyl Alcohol
- Silicon-Carbide Paper
- □ M-Prep Conditioner A
- □ M-Prep Neutralizer 5A
- □ GSP-1 Gauze Sponges
- □ CSP-1 Cotton Applicators
- □ MJG-2 Mylar<sup>®</sup> Tape
- □ TFE-1 Teflon<sup>®</sup> Film
- □ HSC Spring Clamp
- □ GT-14 Pressure Pads and Backup Plates

#### DESCRIPTION

Two-component, solvent-thinned, epoxy-phenolic adhesive for high-performance applications, including high-precision transducers. Solids content 22%. Widest temperature range general-purpose adhesive available. Low viscosity, capable

4 disposable mixing funnels

#### CHARACTERISTICS

Operating Temperature Range:

**Short Term:** -452° to +700°F [-269° to +370°C]. **Long Term:** -452° to +500°F [-269° to +260°C]. **Transducers:** to +450°F [+230°C].

#### Elongation Capabilities:

1% at -452°F [-269°C]; 3% at +75°F [+24°C]; 3% at +500°F [+260°C].

#### Shelf Life:

9 months at +75°F [+24°C]; 15 months at +40°F [+5°C].

**Pot Life:** 6 weeks at +75°F [+24°C]; 12 weeks at +40°F [+5°C].

#### **Clamping Pressure:**

10 to 70psi [70 to 480kN/m<sup>2</sup>]. 30 to 40psi optimum [200 to 275kN/m<sup>2</sup>].

#### **PACKAGING OPTIONS**

#### Kit:

4 bottles [11g ea] Curing Agent4 bottles [14g ea] Resin4 brush caps for dispensing mixed adhesive




Agar Scientific Ltd Unit 7, M11 Business Link Parsonage Lane, Stansted Essex, UK CM24 8GF t: +44 (0)1279 215 506 f: +44 (0)1279 813 105 e: sales@agarscientific.com w: www.agarscientific.com

#### Cure Requirements:

**Recommended Postcure:** 2 hours at 50° to 75°F [30° to

40°C] above maximum operating temperature or cure temperature, whichever is higher.

*High Precision Transducer Postcure:* 2 hours at +400° to +450°F [+205° to +230°C] after wiring.



#### Single Mix Kit:

- 1 bottle [11g ea] Curing Agent
- 1 bottle [14g ea] Resin
- 1 brush cap for dispensing mixed adhesive
- 1 disposable mixing funnel

References: Instruction Bulletin B-130, "Strain Gage Installations with M-Bond 43-B, 600 and 610 Adhesive Systems," included in each kit.

Mylar and Teflon are Registered Trademarks of DuPont

### Notice

Specifications of the products displayed herein are subject to change without notice. Vishay Intertechnology, Inc., or anyone on its behalf, assumes no responsibility or liability for any errors or inaccuracies.

Information contained herein is intended to provide a product description only. No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document. Except as provided in Vishay's terms and conditions of sale for such products, Vishay assumes no liability whatsoever, and disclaims any express or implied warranty, relating to sale and/or use of Vishay products including liability or warranties relating to fitness for a particular purpose, merchantability, or infringement of any patent, copyright, or other intellectual property right.

The products shown herein are not designed for use in medical, life-saving, or life-sustaining applications. Customers using or selling these products for use in such applications do so at their own risk and agree to fully indemnify Vishay for any damages resulting from such improper use or sale.